



OsseoPulse™ AR 300

Dentist Quickstart Guide

1. Set up the OsseoPulse™ prior to patient visit

- Fit hygienic disposables: ear covers, nose piece and array film.
- Position vertical rail with Treatment Array on the appropriate side
- Anterior teeth treatment: use rear adjustment slot - array extends anterior
- Posterior teeth treatment: use inner adjustment slot - array extends posterior

2. Fit the OsseoPulse™ Headset on the patient

- Use Alignment tool: position intra-oral arm on centre of surgical site with gingival margin or alveolar crest aligned with curved arc. Release extra-oral arm and place mark on cheek through perforation.
- Place Headset on patient and adjust earpieces for snug anterior-posterior fit
- Align middle of array window (vertical slot) on mark on cheek, and tighten screws on vertical rail

3. Program the OsseoPulse™

- Ensure the Headset is on the patient correctly
- Turn on device, then press and hold both ▲ and ▼ buttons for 3 seconds
- Choose Prescribe Treatment and press ► button
- Select appropriate treatment and press ► button
- Follow additional instructions
- Complete programming by pressing power button **OFF** and then **ON**

4. Demonstrate the treatment to the patient

- Show patient how to put on Headset and turn on device
- Show patient how to store all parts in carrying case

5. After the patient returns the OsseoPulse™

- Remove and dispose hygienic disposables
- Surface disinfect Headset and Controller with alcohol or surface disinfectant
- Store all parts in carrying case

Resources and instructions available on our website (www.bioluxresearch.com) or in your OsseoPulse™ Dentist User Guide.

Biolux Research Ltd.